# NI 6703/6704 Specifications

This document lists specifications for the NI PCI-6703 and NI PCI/PXI-6704. These specifications are valid for an ambient temperature of 0 to 55  $^{\circ}$ C, unless otherwise noted.

# Analog Output

INL±1 LSB max	Range0.1 to 20.2 mA
Transfer Characteristics	Current Output (NI 6704 Only)
Recommended warm-up time 15 minutes	user-defined values
Resolution 16-bit	Power-on stateIndependent,
(NI 6704 only) 16	Noise100 $\mu V_{rms}$ , DC to 1 MHz
Number of current channels	ProtectionShort-circuit to ground
Number of voltage channels 16	Load capacitance 10,000 pF max

DNL	±1 LSB max
Monotonicity	16 bits, guaranteed

#### **Voltage Output**

Range	±10.1 V
Output coupling	DC
Output impedance	$0.1 \Omega \max$
Current drive	±10 mA max

Range	.0.1 to 20.2 mA
Туре	Source, does not require external excitation source
Output impedance	.1 G $\Omega$ min
Output compliance	.0 to 10 V, not clamped
Noise	.1 $\mu$ A <sub>rms</sub> , DC to 1 MHz
Protection	.Short-circuit and open circuit
Power-up state	.Independent, user-defined values

#### **Accuracy Information**

		Absolute Accuracy				Absolute	Accuracy	
Output	Nominal Range at	% of Reading			Temp Drift	at Full	•	
Туре	Full Scale	24 Hours	90 Days	1 Year	Offset	(%/°C)	24 Hours	1 Year
Voltage	±10.1 V	0.0019%	0.0026%	0.0035%	$\pm 710\mu V$	0.0001%	0.91 mV	1.07 mV
Current*	0.1–20.2 mA	0.0034%	0.0088%	0.0150%	±1,435.0 nA	0.0002%	2.16 µA	4.48 μΑ

Note: Temp drift applies only if ambient is greater than  $\pm 10$  °C of previous external calibration. Absolute Accuracy at Full Scale calculations assume full scale output.

\* NI 6704 only



#### **Dynamic Characteristics**

Settling time (including channel latency)

Accuracy	Time
±0.1%	1.8 ms typ, 5.6 ms max
±0.01%	3.6 ms typ, 11.2 ms max
±0.001%	14.4 ms typ, 48.8 ms max

#### Stability

Offset temperature coefficient

Voltage	5 µV/ ℃
Current (NI 6704 only)	10 nA/ °C

Gain temperature coefficient

Voltage1	ppm/ °C
Current (NI 6704 only)2	ppm/ °C

# Digital I/O

Number of channels8	

Compatibility ......TTL

Power-on state ......Input (high impedance)

Digital logic levels

Level	Min	Max
Input low voltage	_	0.8 V
Input high voltage	2.0 V	—
Output low voltage	_	0.55 V, I <sub>OL</sub> = 16 mA
Output high voltage	2.4 V, I <sub>OH</sub> = 16 mA	—
Input leakage current		10 μΑ

### **Bus Interface**

Type.....Slave

# **Power Requirement**

NI 6703	1.5 A
NI 6704	2.6 A
+12 V	70 mA
-12 V	70 mA <sup>1</sup>

M	Not
$\sim$	pov

**Note** These power usage figures do not include the power used by external devices that are connected to the fused supply present on the I/O connector. They assume that all voltage and current outputs are fully loaded.

Power available at I/O connector.... +4.5 to +5.25 VDC at 0.75 A

#### **Physical**

Dimensions (not including connectors)

NI PCI-6703/6704	$9.9 \times 17.5$ cm
	$(3.9 \times 6.9 \text{ in.})$
NI PXI-6704	10 × 16 cm
	$(3.9 \times 6.3 \text{ in.})$

I/O connector ...... 68-pin male

### **Maximum Working Voltage**

Maximum working voltage refers to the signal voltage plus the common-mode voltage.

Channel-to-earth	11 V, Measurement
	Category I

Measurement Category I is for measurements performed on circuits not directly connected to the electrical distribution system referred to as MAINS voltage. This category is for measurements of voltages from specially protected secondary circuits. Such voltage measurements include signal levels, special equipment, limited-energy parts of equipment, circuits powered by regulated low-voltage sources, and electronics.



**Caution** Do *not* use this device for connection to signals or for measurements within Measurement Categories II, III, or IV.

 $<sup>^1\,</sup>$  NI PXI-6704 devices do not use power from the -12 V rail.

# Environmental

The NI 6703/6704 is intended for indoor use only.

Operating temperature	0 to 55 °C
Storage temperature	–20 to 70 °C
Humidity	5 to 90% RH, noncondensing
Maximum altitude	2,000 m

Pollution Degree ......2



**Note** Clean the device with a soft, non-metallic brush. Make sure that the device is completely dry and free from contaminants before returning it to service.

### Safety

This product is designed to meet the requirements of the following standards of safety for electrical equipment for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA 61010-1

**Note** For UL and other safety certifications, refer to the product label or the *Online Product Certification* section.

# **Electromagnetic Compatibility**

This product meets the requirements of the following EMC standards for electrical equipment for measurement, control, and laboratory use:

- EN 61326 (IEC 61326): Class A emissions; Basic immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- AS/NZS CISPR 11: Group 1, Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



**Note** For the standards applied to assess the EMC of this product, refer to the *Online Product Certification* section.



**Note** For EMC compliance, operate this device with shielded cables.

# CE Compliance $\zeta \in$

This product meets the essential requirements of applicable European Directives as follows:

- 2006/95/EC; Low-Voltage Directive (safety)
- 2004/108/EC; Electromagnetic Compatibility Directive (EMC)

# **Online Product Certification**

Refer to the product Declaration of Conformity (DoC) for additional regulatory compliance information. To obtain product certifications and the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

# **Environmental Management**

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *NI and the Environment* Web page at ni.com/environment. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

#### Waste Electrical and Electronic Equipment (WEEE)



**EU Customers** At the end of the product life cycle, all products *must* be sent to a WEEE recycling center. For more information about WEEE recycling centers, National Instruments WEEE initiatives, and compliance with WEEE Directive 2002/96/EC on Waste and Electronic Equipment, Visit ni.com/environment/weee.

#### 电子信息产品污染控制管理办法 (中国 RoHS)

•

中国客户 National Instruments 符合中国电子信 息产品中限制使用某些有害物质指令 (RoHS)。关于 National Instruments 中国 RoHS 合规性信息, 请登录 ni.com/environment/rohs\_china。 (For information about China RoHS compliance, go to ni.com/environment/rohs\_china.)

#### **Device Pinouts**

	$\frown$				(	$\frown$		
AO 0 (V)	34	68	AO GND 0		AO 0 (V)	34	68	AO GND 0/16
AO GND 1	33	67	NC		AO GND 1/17	33	67	AO 16 (I)
NC	32	66	AO 1 (V)		AO 17 (I)	32	66	AO 1 (V)
AO 2 (V)	31	65	AO GND 2		AO 2 (V)	31	65	AO GND 2/18
AO GND 3	30	64	NC		AO GND 3/19	30	64	AO 18 (I)
NC	29	63	AO 3 (V)		AO 19 (I)	29	63	AO 3 (V)
AO 4 (V)	28	62	AO GND 4		AO 4 (V)	28	62	AO GND 4/20
AO GND 5	27	61	NC		AO GND 5/21	27	61	AO 20 (I)
NC	26	60	AO 5 (V)		AO 21 (I)	26	60	AO 5 (V)
A0 6 (V)	25	59	AO GND 6		A0 6 (V)	25	59	AO GND 6/22
AO GND 7	24	58	NC		AO GND 7/23	24	58	AO 22 (I)
NC	23	57	AO 7 (V)		AO 23 (I)	23	57	AO 7 (V)
AO 8 (V)	22	56	AO GND <sup>1</sup>		AO 8 (V)	22	56	AO GND <sup>1</sup>
NC	21	55	AO GND 8		AO 24 (I)	21	55	AO GND 8/24
AO GND 9	20	54	AO 9 (V)		AO GND 9/25	20	54	AO 9 (V)
AO GND <sup>1</sup>	19	53	NC		AO GND <sup>1</sup>	19	53	AO 25 (I)
AO GND 10	18	52	AO 10 (V)		AO GND 10/26	18	52	AO 10 (V)
AO 11 (V)	17	51	NC		AO 11 (V)	17	51	AO 26 (I)
NC	16	50	AO GND 11		AO 27 (I)	16	50	AO GND 11/27
AO 12 (V)	15	49	AO GND 12		AO 12 (V)	15	49	AO GND 12/28
AO GND 13	14	48	NC		AO GND 13/29	14	48	AO 28 (I)
NC	13	47	AO 13 (V)		AO 29 (I)	13	47	AO 13 (V)
AO 14 (V)	12	46	AO GND 14		AO 14 (V)	12	46	AO GND 14/30
AO GND 15	11	45	NC		AO GND 15/31	11	45	AO 30 (I)
NC	10	44	AO 15 (V)		AO 31 (I)	10	44	AO 15 (V)
P0.7	9	43	AO GND <sup>1</sup>		P0.7	9	43	AO GND <sup>1</sup>
P0.6	8	42	D GND		P0.6	8	42	D GND
P0.5	7	41	D GND		P0.5	7	41	D GND
P0.4	6	40	RESERVED		P0.4	6	40	RESERVED
P0.3	5	39	D GND		P0.3	5	39	D GND
P0.2	4	38	RESERVED		P0.2	4	38	RESERVED
P0.1	3	37	D GND		P0.1	3	37	D GND
P0.0	2	36	D GND		P0.0	2	36	D GND
+5 V	1	35	D GND <sup>1</sup>		+5 V	1	35	D GND <sup>1</sup>
V = Voltage V = Voltage								
I = Current								
NC = No Connect NC = No Connect								
1 No Connect when using the SH68-68-D1 cable. 1 No Connect when using the SH68-68-D1 cable.								
Figure 3. NI 6703	3 Con	necto	r Pin Assignments		Figure 4. NI 6704	4 Con	necto	r Pin Assignments



LabVIEW, National Instruments, NI, ni.com, the National Instruments corporate logo, and the Eagle logo are trademarks of National Instruments Corporation. Refer to the Trademark Information at ni.com/trademarks for other National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: Help»Patents in your software, the patents.txt file on your media, or the National Instruments Patent Notice at ni.com/patents. Refer to the Export Compliance Information at ni.com/legal/ export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data.

© 2009–2012 National Instruments Corporation. All rights reserved.